



America

# C E R T I F I C A T E

No. U8 095184 0022 Rev. 00

**Holder of Certificate:** **E. ZOLLER GmbH & Co. KG**

Gottlieb-Daimler-Str. 19  
74385 Pleidelsheim  
GERMANY

**Production** 095184

**Facility(ies):**

**Certification Mark:**



**Product:** **Electronic measuring equipment**

**Model(s):**

**SMILEyX-z, SMILECPT-z, SMILEyT-z**  
y: can be 320, 420 or 620 and specified the height of the  
z- axis  
z: is used as serial number and can be 00050 to 99999

**Parameters:**

Rated voltage: 100-120 / 200-240 V  
Rated frequency: 50/60 Hz  
Rated power: 600 VA  
Protection class: I  
Air pressure: 6-8 bar

**Tested**

UL 61010-1:2012/R:2018-11

**according to:**

CAN/CSA-C22.2 No. 61010-1:2012/A1:2018-11

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited Certification body.

**Test report no.:** 028-713171751-000

**Date,** 2020-04-03

( Benedikt Pulver )